

August 9, 2020

Three Forks Wastewater Testing Results

Result Summary: Three Forks Wastewater Sample was positive but low.

Sample Description:

- 1) A composite sample of wastewater (1.0 L total) inflow to the Three Forks treatment plant was captured on 8/5/2020 using an auto-sampler over the previous 24-hour period. Referred to below as “Inflow” samples in **Table 1**.

Testing Information and Raw Data:

Testing for the presence and abundance of the SARS-CoV2 genome in the above samples was performed using a kit designed by the US Centers for Disease Control and Prevention (CDC 2019-Novel Coronavirus (2019-nCoV), Real-Time RT-PCR Diagnostic Panel). Importantly, this test kit was originally designed to detect the virus in human samples and NOT wastewater or other kinds of environmental samples. The test was used here to determine whether a detectable amount of virus was present. Results need to be interpreted with caution, as described below.

Each of the above samples were split and processed as two replicates. Two tests were performed on each replicate and two independent locations on the SARS-CoV2 genome were targeted (N1 and N2). RNA was isolated from inactivated/concentrated samples, reverse-transcribed to DNA and used as template in quantitative PCR reactions as per kit instructions. Results were recorded as cycle threshold (Ct) numbers based on test interpretation guidelines described by the CDC. A standard curve was generated using a pre-made virus target and used to calculate the number of genomes in each sample.

Results were as follows:

Three Forks Sample ID	Replicate ID	Target	Ct	Potential Genomes per liter
Inflow_1	N1.1	N1	40.7964	207
Inflow_1	N1.2	N1	41.3636	130
Inflow_1	N2.1	N2	44.0652	14
Inflow_1	N2.2	N2	undetermined	0
Inflow_2	N1.1	N1	40.8922	191
Inflow_2	N1.2	N1	41.7912	92
Inflow_2	N2.1	N2	44.0652	14
Inflow_2	N2.2	N2	41.2066	148

Interpretation:

Signal in most replicates was above our limit of detection in the Inflow samples. Based on our experience with wastewater testing, this is evidence of low levels of virus in wastewater. Levels of virus in this sample was lower than the previous week’s sample (taken 7/29/2020).

Relevant text from CDC guidelines:

“...a specimen is considered positive for 2019-nCoV if all 2019-nCoV marker (N1, N2) cycle threshold growth curves cross the threshold line within 40.00 cycles (< 40.00 Ct).”

“When all controls exhibit the expected performance and the cycle threshold growth curve for any one marker (N1 or N2 but not both markers) crosses the threshold line within 40.00 cycles (< 40.00 Ct) the result is inconclusive.”